



REV	DATE	DESCRIPTION
14	5/14/98	RADIAL CONVERSION PER E970587 & 980015
13	09/10/97	CHANGED ERROR TRANSFORMER WIRING ON SHEET 3 OF 3 PER ECO #970400 PER E98048.
12	11/13/96	REV CHG TO MATCH PICT
11	NOT USED	
10	09/10/96	ADDED GOOD TRAINING TO PICTORIAL (SHT 3 OF 3) PER ECO #960554.
9	08/14/96	REV CHG TO REFLECT PIC PER CO
8	05/20/96	LMA
5	01/03/95	ENW
4	05/16/94	LMA
3		
2	11/23/93	ENW
1	09/03/93	SWR

SIGNATURES	DATE	CHK'D
1880 BORMAN DR ST. LOUIS, MISSOURI 63148	04/05/93	
PROJECT NAME:	SVT-3 PRO	
DRAWING NO.:	07S426-01	
REV.:	REV. 15	
SCALE:	1:1	
FILE NAME:	42601HN	
DATE:	07/28/98	
PLT TIME:	09:18:47	
ADD'D:	ADDED BLUE/WHIT WIRE TO XFRMR PER E98021HN	

- NOTES**
- 1) CAUTION: SHOCK HAZARD! THIS UNIT CONTAINS HAZARDOUS VOLTAGE. DISCONNECT POWER AND BE SURE POWER SUPPLY IS DISCHARGED BEFORE TOUCHING INTERNAL PARTS.
 - 2) UNLESS NOTED, RESISTOR VALUES IN OHMS, 1/4W-5% TOL. CAPACITOR VALUES IN MICROFARADS, 50V-10% TOL.
 - 3) VOLTAGES ARE MEASURED WITH 1 MEGOHM OSCILLOSCOPE AND 10 MEGOHM DIGITAL VOLTMETER.
 - 4) CHASSIS GROUND ∇ DIRTY GROUND ∇ SIGNAL GROUND ∇

CAUTION: THIS SCHEMATIC IS PROVIDED FOR USE BY QUALIFIED PERSONNEL TO AVOID RISK OF ELECTRIC SHOCK. REFER SERVICING TO QUALIFIED SERVICE PERSONNEL. DO NOT PERFORM ANY SERVICING BEYOND THAT EXPLAINED IN THE OPERATING INSTRUCTIONS.